Revision 1
Issue 2

EC DECLARATION OF CONFORMITY

This is to declare that the:

EN 13348 Degreased Medical Copper Tube

Product Code:
10CP
12CP
15CP
22CP
28CP
35CP
42CP
54CP
76CP
108CP

Manufactured under BS EN 9001 & BS EN 13485 Medical Devices: Quality Management Systems and supplied by:

Precision UK Ltd
CPX Technology Building, Pepper Road, Hazel Grove
Stockport, SK7 5BW, UK

Tel: 00 44 (0) 161 487 2822
Fax: 00 44 (0) 161 487 2816

info@precisionuk.co.uk

www.precisionuk.co.uk

Detailed hereon has been manufactured, inspected and tested in accordance with requirements of the purchase order and complies with the requirements of:
HTM 2022, HTM 02-01, C11, BS EN 13348, BS EN 1057, BS 1710, BS 2871 Part 1 Table X, BS EN 7396-1 BS EN 723, BS EN 7396-2 and CE marked under annex II of the Medical Device Directive 93/42/EEC (CE0086).
Are class IIa Medical Devices as defined by Council Directive 93/42/EEC
Under our appointed Notified Body BSI notified body no. 0086
Chemical composition CU.DHP to ISO 1190-1/CW024A to EN 1412
Each pipe is cleaned to have a maximum carbon level 0.20mg/dm

All Degreased Medical Gas Copper Tube is batch numbered for traceability prior to packing and delivery

Signed: K Pugh

Name: KEVIN PUGH

Position: TECHNICAL MANAGER

Date: 04/11/2015